

What are the inter-instrument agreement and repeatability specifications for SpectraVision?

Inter-Instrument Agreement : dE^* 0.15 Avg and 0.25 Max on 12 BCRA tiles

Repeatability: dE^* 0.03 Max on White BCRA tile

Author: Datacolor

Saved From: <http://knowledgebase.datacolor.com/article-708.html>